

Reduce waste. Go green.

A reprocessing solution designed to:

- **Meet sustainability goals**
- **Reduce hospital waste**

Kendall DL™ Disposable Cable and Lead Wires, Kendall SCD™ Sequential Compression Sleeves and Nellcor™ Pulse Oximetry Sensors are intended for single-patient use.

Place the used sensors, compression sleeves, cables and lead wires in the provided collection hampers after the patient is discharged.

Kendall DL™ Disposable Cable and Lead Wires
After every patient use:



Detach all electrodes by pushing the pink button and releasing the head of the lead set from the electrode snap.



Disconnect the cable or lead set from the bedside monitor or telemetry box.



Leave the reusable adapter connected to the bedside monitor or telemetry box.

Kendall SCD™ Sequential Compression Sleeves
After every patient use:



Detach tubing from the sequential compression sleeves.



Leave tubing connected to the controller.



Deposit used products in the hampers we have provided for reprocessing.



Do not use collection bins to discard the following items:



Sharps

No scalpels, needles, syringes, sutures or glass



Waste

No personal equipment, trash, patient care disposables or linens



Batteries

Remove from device but do not place in collection container



Kendall SCD™ controllers

Leave tubing connected to the Kendall SCD™ controllers



Kendall SCD™ tubing

Leave tubing connected to the Kendall SCD™ controllers



Reusable Kendall DL™ adapters

Leave adapters connected to bedside monitors or telemetry boxes

For more information call **844.268.2583**
or email **gogreen@cardinalhealth.com**